

Quantitative determination of Lactate Dehydrogenase (LDH) in serum/plasma
Only for *In Vitro* Diagnostic use

ORDER INFORMATION

REF	Cont.
LDHSLR 10	1 X 10 mL
LDHSLR 25	1 X 25 mL

CLINICAL SIGNIFICANCE

Lactate dehydrogenase (LDH) is an enzyme with wide tissue distribution in the body. Increased levels of the enzyme are found in serum in liver disease, myocardial infarction, renal disease, muscular dystrophy and anemia.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

Method

Optimized UV-test according to IFCC (International Federation of Clinical Chemistry and Laboratory Medicine) and DGKC (German Society of Clinical Chemistry)

PRINCIPLE

Lactate is oxidised to Pyruvate in the presence of NAD by the action of lactate dehydrogenase. The rate of formation of NADH is directly proportional to LDH concentration.

REAGENT

Reagent I : Buffer reagent
Reagent II : Substrate reagent

REAGENT PREPARATION

Mix 9 ml of Buffer reagent with 1 ml of Substrate reagent.

REAGENT STORAGE AND STABILITY

Prior to use:

When stored between 2-8°C the reagent is stable until the expiration date stated on the bottle and kit box label.

Reconstituted Reagent:

When stored capped at 2-8°C, the reagent is stable for at least 7 days.

WARNING AND PRECAUTIONS

- For in vitro diagnostic use.
- Do not use components beyond the expiration date.
- Do not mix materials from different kit lot numbers.
- Exercise the normal precautions required for handling all laboratory reagents.
- The reagent contains preservative. Do not swallow. Avoid contact with skin and mucous membranes.
- For detailed information refer Material Safety Data Sheet.

WASTE MANAGEMENT

Please refer to local legal requirements.

MATERIALS REQUIRED BUT NOT PROVIDED

- NaCl solution 9 g/L
- General laboratory equipment

SAMPLE COLLECTION AND PRESERVATION

Serum or heparin plasma or EDTA plasma

Separate at the latest 1h after blood collection from cellular contents.
1 month at 4 – 8°C

Discard contaminated specimens!

ASSAY PROCEDURE

Operating Instructions

- Check reagent inventories at least daily to ensure that quantities are sufficient for the planned work load.
- Bring all reagents, standard and samples to room temperature 18 - 28°C, prior to analysis.

AUTOMATED PARAMETERS	
Wavelength	340nm
Cuvette	1 cm light path
Reaction Temperature	37°C
Measurement	Against distilled water
Reaction Type	Kinetic test
Reaction direction	Increasing
Sample Volume	25 µl
Reagent Volume	1000 µl
Delay/lag/time	60 Secs
Interval Time	30 Secs
No. of Readings	04
Blank Absorbance limit	≤ 0.6
Factor	6592
Low Normal at 37°C	80 IU/L
High Normal at 37°C	285 IU/L
Linearity	1200 IU/L

MANUAL ASSAY PROCEDURE

Pipette into Test Tubes

Working Reagent	1000 µl
Sample	25 µl

Mix well and after 1 min incubation, measure the change in absorbance per min. ($\Delta A/\text{min.}$) for next 2 minutes.

SAMPLE DILUTIONS

- This method is linear upto a concentration of 1200 IU/L.
- Dilute samples above this concentration 1:1 with 0.9% saline
- Repeat assay. Multiply the result by 2.

CALCULATION

Results are calculated, usually automatically by the instrument, as follows:

$$\text{LDH Activity (IU/L)} = \Delta A/\text{min} \times 6592$$

CALIBRATORS AND CONTROLS

For the calibration of automated photometric systems the commercially available suitable multi-calibrator is recommended.

This method has been standardized against the original IFCC formulation.

It is recommended to run a normal and a pathological control serum which is commercially available to verify the performance of the measured procedure. The value of controls should fall within the established limit.

Each laboratory should establish corrective action in case of deviations in control recovery.

PERFORMANCE CHARACTERISTICS







WITHIN RUN

Sample	Mean Concentration	SD	CV %
Level 1	121.74	3.81	3.13%
Level 2	402.31	3.14	1.03%

RUN TO RUN

Sample	Mean Concentration	SD	CV %
Level 1	120.70	2.98	2.46%
Level 2	404.31	3.81	0.94%

GLOSSARY OF SYMBOL

	Consult Instruction for Use	LOT	Lot Number
REF	Catalog Number		Date of Manufacturing
	Store between		Use By or Expiration Date
	Manufacturer	IVD	For <i>in vitro</i> Diagnostic use only
	Keep away from sunlight	CONT	Content of the kit



LAB-CARE DIAGNOSTICS (INDIA) PVT. LTD.
C1 Type, Shed No.: 3225, Chemical Zone,
GIDC Sarigam – 396155, Dist. Valsad, Gujarat, India.
Tel.: +91 22 2554 2109 /1558
Email: accucare@labcarediagnostics.com;
Website: www.labcarediagnostics.com

LINEARITY

The method is linear upto a concentration of 1200 IU/L. Dilute samples above this concentration 1:1 with 0.9% saline solution and repeat assay. Multiply the result by 2.

Limit of detection: The limit of detection for LDH is 5 IU/L.

METHOD COMPARISON

A comparison of Accucare LDH with a commercially available assay (x) using 20 samples gave following results: $R^2 = 0.9800$

REFERENCE VALUES

Males	30°C	50 - 166 IU/L
	37°C	80 - 285 IU/L
Females	30°C	60 - 132 IU/L
	37°C	103 - 227 IU/L

The reference values are to be considered as indicative only. Every laboratory should establish its own normal range.

LIMITATION OF THE PROCEDURE

- For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

INTERFERENCE

- Ascorbic Acid: No interference found upto 30 mg/dL.
- Bilirubin: No interference found upto 40mg /dL.
- Lipemia: No interference found upto 1500 mg/dL.
- These characteristics have been obtained using an automatic analyzer. Results may vary if a different instrument or a manual procedure is used.

BIBLIOGRAPHY

- Thomas L. Clinical laboratory diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft;1998. 89–94.
- Moss DW, Henderson AR. Clinical enzymology In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry 3rd ed. Philadelphia: W.B Saunders Company;1999.617–721.
- Schumann G, Bonora R, Ceriotti F, Féraud G et al. IFCC primary reference procedure for the measurement of catalytic activity concentrations of enzymes at 37 °C. Part 3: Reference procedure for the measurement of catalytic concentration of lactate dehydrogenase. Clin Chem Lab Med 2002;40:643-48.
- Soldin JS, Hicks JM. Pediatric reference ranges. Washington: AACCC Press:1995:95. 8. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanisms, detection and prevention. Clin Chem Lab med 2007; 45(9): 1240–1243.
- Deutsche Gesellschaft für Klinische Chemie. (German Society for Clinical Chemistry). Recommendation for the determination of the catalytic concentration of lactate dehydrogenase at 37 °C. Eur J Clin Chem Clin Biochem 1993;31:897-9.